



DESNA CONTROLS & WEIGHING INC.
 412 MANITOU ROAD SE, CALGARY, AB T2G 4C4
 Phone: 403/243-4848 Fax: 403/243-9338

MC Accreditation Certificate # A0258

SCALE INSPECTION REPORT:

CUSTOMER: Customer Name
ADDRESS: Customer Address

DATE INSPECTED: July 22, 2015
TECHNICIAN: Joe S.

NEXT INSPECTION DUE: October 22, 2015

This is to certify that the following weighing equipment has been verified/calibrated by Desna Controls & Weighing Inc. and the device has been found to provide a readout accurate to within +/- 0.1% of the indicated capacity unless otherwise noted.* This system was checked using government approved test weights with verification traceable to the NIST in accordance with ASTM E4-99.

INDICATOR: RL 920i **SERIAL NO:** 218175
BASE: RL Roughdeck 4 ft x 4 ft **SERIAL NO:** 632514-5
CAPACITY: 2,500 kg x 0.5 kg **LEGAL FOR TRADE:** NO

CORNER TEST:

1	2
5	
4	3

Front

Total Test Weight Applied
500 kg

AS FOUND				AS LEFT		
Corner	Applied Weight	Scale Reading	Scale Error	Applied Weight	Scale Reading	Scale Error
1	500 kg	498 kg	-2 kg	500 kg	500 kg	0 kg
2	500 kg	499 kg	-1 kg	500 kg	500 kg	0 kg
3	500 kg	499 kg	-1 kg	500 kg	500 kg	0 kg
4	500 kg	497 kg	-3 kg	500 kg	500 kg	0 kg
5	500 kg	499 kg	-1 kg	500 kg	500 kg	0 kg
Span	Applied Weight	Scale Reading	Scale Error	Applied Weight	Scale Reading	Scale Error
	0 kg	0 kg	0 kg	0 kg	0 kg	0 kg
	500 kg	499 kg	-1 kg	500 kg	500 kg	0 kg
	1000 kg	999 kg	-1 kg	1000 kg	1000 kg	0 kg
	1500 kg	1498 kg	-2 kg	1500 kg	1500 kg	0 kg
	2000 kg	1998 kg	-2 kg	2000 kg	2000 kg	0 kg
	2500 kg	2497 kg	-3 kg	2500 kg	2500 kg	0 kg
	2000 kg	1998 kg	-2 kg	2000 kg	2000 kg	0 kg
	1500 kg	1498 kg	-2 kg	1500 kg	1500 kg	0 kg
	1000 kg	999 kg	-1 kg	1000 kg	1000 kg	0 kg
	500 kg	499 kg	-1 kg	500 kg	500 kg	0 kg
	0 kg	0 kg	0 kg	0 kg	0 kg	0 kg

Weights Used: 5 x 500kg S/N: DC101 - DC105

COMMENTS: Tested to capacity. Corners adjusted Scale recalibrated. Reading accurate and stable.

NOTES: Quarterly Preventative Maintenance Inspection Schedule

Calibration sticker applied to indicator by technician.